



ES CERTIFIKAT o o d o b r i t v i t i p a m e r i l a

EC type approval certificate

št. /No. SI 13-05-012

Neavtomatska tehnica, tip **MJ yyy, MPE yyy, MS yyy**
Non-automatic weighing instrument, type MJ yyy, MPE yyy, MS yyy

Certifikat izdal
Issued by

Urad RS za meroslovje
Metrology Institute of the Republic of Slovenia (MIRS)
Tkalska ulica 15
SI-3000 Celje, Slovenija

Priglašen organ št. 1376
Notified body No.

na podlagi
in accordance with

1. odstavka 9. člena Zakona o meroslovju (Ur. list RS, št. 26/05) in 17. člena Pravilnika o meroslovnih zahtevah za neavtomatske tehnice (Ur. list RS, št. 97/03), ki povzema vsebino Direktive 2009/23/ES Evropskega parlamenta in Sveta (kodificirana različica Direktive Sveta 90/384/EGS).
Act on metrology (OG of RS, No. 26/05), Article 9, Paragraph 1, and Rules on metrological requirements for NAWI (OG of RS, No. 97/03), Article 17, which implement Directive 2009/23/EC of the European Parliament and of the Council (codified version of Council Directive 90/384/EEC).

izdano vložniku
issued to

Vage d.o.o.
Koledovčina 2a
HR - 10000 Zagreb, Croatia

za
in respect of

Neavtomatska elektronska tehnica, eno ali več razdelčna.
Non-automatic electronic weighing instrument. single-range or multiple-range, single-interval or multi-interval.

Točnostni razred **(III)** or **(III)**
Accuracy class

Največje število preskusnih razdelkov:
The maximum number of verification scale intervals:

$n \leq 10000$ za razred **(III)**, tehnice z enim razdelkom
*for Class **(III)** instruments with single interval*
 $n \leq 3000$ za razred **(III)**, večrazdelčne tehnice
*for Class **(III)** instruments with multi-interval operation*
 $n \leq 1000$ za razred **(III)**, tehnice z enim razdelkom
*for Class **(III)** instruments with single interval*

veljavnost do
valid until

30. 9. 2023

Glavne značilnosti, pogoji in omejitve so podani v prilogi, ki je sestavni del tega certifikata in obsega 10 strani.

The principal characteristics, approval conditions and special conditions, if any, are set out in the Appendix hereto, which forms part of the approval documents and consists of 10 pages.

Štev. certifikata: 6413-4/2013/2
Ljubljana, 30. 9. 2013

Postopek vodil:

mag. Gašper Vindišar
podsekretar



dr. Samo Kopač
direktor